



**National Patent
Analytical Systems, Inc.**

On this date 10/28/11, NPAS received instrument # 104709.

The instrument was sent from the Vermont Department of Health (VDH) so that NPAS could install updated software as well as inspect the instrument. At NPAS' request VDH provided copies of the Technical Support Inquiries (TSI) regarding the instrument. The calibration history of the instrument remained in the instrument so that NPAS could review its history.

NPAS examined the internal configuration of the instrument and did / did not find physical manipulation or tampering of the components located within the instrument itself.

NPAS did performance and accuracy checks on a certified 0.100 ethanol standard. NPAS found that the instrument passed / failed the performance and accuracy checks.

NPAS reviewed the instrument as to its DataMaster DMT Inspection, QA and Conformance Testing Procedure and the instrument performed / did not perform within its procedure.

NPAS examined the calibration records and did / did not observe anomalies.

NPAS testing indicates that there is no reason why this instrument would not have performed accurately and reliably in the field.

*NPAS
11-14-11*

Print Name of Evaluator: Michael J. Barbely Jr.

Signature of Evaluator and Date: Michael J. Barbely Jr. 10/31/11

Print Name of Reviewer: DAVID RADOMSKI

Signature of Reviewer and Date: David Radowski 11/7/11

Packing List

From: National Patent Analytical Systems Inc.
2090 Harrington Mem. Rd., Mansfield, Ohio 44903

Cust No. VT1678J Ord Date 10/28/11
Sales Order No. 44926

Ship Date 11/8/11
Packing List No. 69196

Cust Purch Ord No. KIRK KIMBALL

Ship To:

Vermont Dept of Health

195 Colchester Avenue

Burlington Vt 05401

Bill To:

Vermont Dept of Health

P O Box 70

Burlington

Vt 05402 Freight

25.00

Attn:

Ship Via FedEx

Comments Service for Instrument Serial Number 104709

Inv Part No	Quantity Ordered	This Shipment	Back Ordered	Total Ship To Date	Description	Customer Part No.	Drawing Part No. / Rev
9DMT	1.0	1		1.0	DATAMASTER DMT SERVICE RETURN		
S/N 104709						None Given	No Draw # / NONE
Serial Nos.	104709						

AK-SC
11-14-11